


送樣單位: 春星公司
 樣品名稱: 經 "Mini Kiko" 處理的自來水
 半高寬: 77Hz



Current Data Parameters
 NAME servol7chnumxing
 EXNO 5
 PROCNO 1

F2 - Acquisition Parameters

Date_ 20111222
 Time 16.06
 INSTRUM dmx500
 PROBD 5 mm TBI 1H-BB
 PULPROG ZF
 TD 512
 SOLVENT CDCl3
 NS 2000
 DS 0
 SWH 4006.410 Hz
 FIDRES 7.825020 Hz
 AQ 0.0640724 sec
 RG 20480
 DW 124.800 usec
 DE 20.00 usec
 TE 0.0 K
 D1 0.50000000 sec
 MCREST 0.00000000 sec
 MCWPK 0.01500000 sec

==== CHANNEL F1 =====

NUC1 17O
 P1 20.10 usec
 PL1 6.00 dB
 SFO1 67.8000675 MHz

F2 - Processing Parameters

SI 32768
 SF 67.8000909 MHz
 WDW no
 SSB 0
 LB 0.00 Hz
 GB 0
 PC 4.00